

# Performance Testing / Calibration Testing for Thermo EPD's and JCS ED3 Real Time Dosimeters

We are now offering, In partnership with John Caunt Scientific, an express annual performance test for the ED3 Extremity Electronic Dosimeter Probes.




ED3 Detector Probes should be checked annually to ensure their response to radiation is within specification. For the D1 Probe angular response is checked to confirm the integrity of the internal filters.

Phoenix Dosimetry in partnership with Thermo Fisher Scientific now offer an express calibration service for the EPD-2 Personal Electronic Dosimeters.



Our laboratory is now fully equipped with the essential Thermo Fisher calibration and test equipment which is specifically designed for the EPD-2 units.



To use this service simply email us at [support@phoenix-dosimetry.co.uk](mailto:support@phoenix-dosimetry.co.uk) and let us know how many and which type of units you would like us to test / cross calibrate

**Phoenix Dosimetry Ltd** Unit 6, Lakeside Business Park, South Lane, Southport, Merseyside, CH43 9DN

**Radiological Performance Certificate of Test No: 1000**

**Details of instrument and use:**  
 Description: ED3 ED3 with 01 Detector  
 Serial No: 00001438  
 Use: Measurement of X and Y radiation dose at and above 60 kR/h

**Details of Test:**  
 Date: 4<sup>th</sup> July 2017  
 Type of Test in bold: **Test Before First Use**  
 Test Conditions: The instruments dose response was tested together with the polar response which was determined at 90 and 180° in the horizontal plane. All measurements were performed on an X-Ray set fitted with Tungsten filter at 90kVp. The dosimeter was mounted lengthwise on the front face of an ISO PMMA rod phantom with a 25 mm diameter and 300 mm length.

**Test reference point:** Front of phantom  
**Orientation:** Radiation beam normal to detector reference mark, unless angle to beam noted  
**Uncertainty:** The uncertainties associated with this measurement are not expected to be above 2.5% at the 95% confidence level.  
**Traceability:** The measurements were performed against a reference ED-3 Dosimeter directly calibrated at the PH / UKAS facility.

**Results of Test:** Satisfactory for measurements performed. See over.

Compliance with the Ionising Radiations Regulations 1989:  
 These results are typical of type and the instrument / detector is suitable for the use described in Section 6, under the terms of the Regulations and the associated Approved Code of Practice.

**Test Performed By:** Ken Bosley **Signature:**  
**Date of Issue:** 11<sup>th</sup> July 2017

Telephone: +44 (0)1524 871590 Website: www.phoenix-dosimetry.co.uk Email: sales@phoenix-dosimetry.co.uk  
 Phoenix Dosimetry is a trading name of Phoenix Dosimetry Limited. Registered in England. No. 08429426

**Sample Test Certificate**

**Phoenix Dosimetry Ltd** Certificate of Performance Verification Reference: Report From: 01/03/2017 Report No: 21/03/2017

Thermo Scientific Electronic Personal Dosimeter (EPD) M1.2  
 Temperature: 22 °C Ambient Bar Number: 10  
 Humidity: 50% RH 46-50 °C/106 11584 11584  
 The stability checks for EPDs are performed against approximations at NIST traceability.

**Performance Verification tolerance for EpDose is:** 10% of the reference standard  
**Performance Verification tolerance for Dose is:** 20% of the reference standard

1. Does include: High Temp. Test, Source, Ion Beam, or any other tests in testing reports. See: sales@phoenix-dosimetry.co.uk  
 2. Does not include: requests for heavy metals (Dose and/or DoseRate)

EpDose (DoseRate)	EpDose (Dose)	High Temp. Test	Source	Ion Beam	Other Tests	Test Results	Pass/Fail
0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	Pass
0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	Pass
0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	Pass
0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	Pass
0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	Pass
0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	Pass
0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	Pass
0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	Pass
0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	Pass
0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	0.000000	Pass



Turnaround times are typically 1 week. Please email us now for any further details at [support@phoenix-dosimetry.co.uk](mailto:support@phoenix-dosimetry.co.uk)